



# Certificate of Analysis

## Certified Reference Material

---

### FRUT-1

#### Isotopic Certified Reference Material of fructose

FRUT-1 is a fructose Certified Reference Material (CRM) that is certified for carbon isotope delta values. FRUT-1 is part of a suite of sugar based isotopic CRMs from the National Research Council of Canada (NRC). Each unit of FRUT-1 contains approximately 1 g of fructose. In conjunction with BEET-1 and GALT-1, this CRM is intended for the calibration of procedures and development of methods for normalizing carbon isotope delta measurements of sugar-based or other carbon-bearing materials to the Vienna Peedee Belemnite (VPDB) scale.

The following table shows the certified carbon isotope delta value that has been established for this fructose isotopic CRM. The combined expanded uncertainty ( $U_{CRM}$ ) in the certified value is equal to  $U_{CRM} = k u_c$  where  $u_c$  is the combined standard uncertainty calculated according to the JCGM Guide [1] and  $k$  is the coverage factor. A coverage factor of two (2) was applied. It is intended that  $U_{CRM}$  accounts for every aspect that reasonably contributes to the uncertainty of the measurement. The certified carbon isotope delta value and its expanded uncertainty are expressed on the VPDB scale with a value of  $-46.6\text{‰}$  assigned to LSVEC and  $+1.95\text{‰}$  to NBS19 [2].

**Table 1: Certified quantity value and uncertainty for FRUT-1**

Material	Isotope delta, $\delta_{VPDB}^{(13C)}$ , ‰	Combined standard uncertainty, ‰	Combined expanded uncertainty, ‰
FRUT-1	-10.95	0.03	0.06

#### Certified values

Certified values are considered to be those for which the NRC has the highest confidence in accuracy, and that all known and suspected sources of bias have been taken into account and are reflected in the stated expanded uncertainties. Certified values are the best estimate of the true value and uncertainty.

The certified value for carbon isotope delta in FRUT-1 was determined by combining elemental analysis with isotope ratio mass spectrometry (EA-IRMS) measurement results from seven participating expert laboratories (including the NRC) using a random laboratory effects statistical model.

### Intended use

This certified reference material is primarily intended to be used with at least one other carbon reference material [3], preferably as a set of three with BEET-1 [4] and GALT-1 [5] CRMs, for normalizing carbon isotope delta measurements of sugar-based or other carbon-bearing materials to the VPDB scale. A minimum sample mass of 550 µg is recommended.

### Storage and sampling

Unopened ampules of FRUT-1 may be stored at room temperature in a clean location out of direct sunlight. Once the ampule is opened, it is recommended that all contents be immediately transferred to a clean, dry vial with a tight-fitting closure. The vial should be stored at room temperature in a dry box or desiccator out of direct sunlight when not in use.

### Preparation of material

FRUT-1 was prepared from a lot of D-fructose. Fructose was sieved in small aliquots using 0.3 and 0.4 mm mesh sieves. All fructose between 0.3 and 0.4 mm were collected into 6 plastic bottles pre-flushed with argon, capped, then placed in a –80 °C freezer for two hours. Uncapped bottles of fructose were then freeze-dried at –50 °C under vacuum for two days. Freeze-dried fructose was stored in a desiccator until packaged. Prior to packaging, each of the 6 plastic bottles were tumbled for further homogenization. A unit of FRUT-1 comprises a sealed, argon-flushed ampule containing approximately 1 g of fructose [6].

### Stability

FRUT-1 is isotopically stable when stored under recommended conditions and typical transportation temperatures. The transportation and long-term stability were assessed at the NRC using EA-IRMS. The transportation stability was carried out via an isochronous approach, over a period of two weeks, using unopened ampoules at temperatures ranging from –20 °C to +37 °C. The long-term stability was assessed at room temperature for a period of 5 years.

### Homogeneity

The homogeneity of the carbon isotope delta values in FRUT-1 was assessed by analyzing 31 units of FRUT-1. These results were evaluated using Bayesian random effects model and were included in the calculation of the certified value.

### Uncertainty

Included in the combined standard uncertainty estimate ( $u_c$ ) are the uncertainties from characterization by the NRC and six external laboratories ( $u_{char}$ ), and the uncertainty related to between-bottle variation ( $u_{hom}$ ). Expressed as standard uncertainties, these components are listed in Table 2.

**Table 2: Uncertainty components for FRUT-1**

Material	$u_c$	$u_{char}$	$u_{hom}$
FRUT-1	0.03	0.021	0.020

The uncertainty due to characterization ( $u_{char}$ ) includes the uncertainty due to measurement of the sugar samples and the international reference materials, uncertainty due to possible

inconsistencies between the laboratory results, and uncertainty assigned to the values of the international reference materials.

### **Metrological traceability**

The *Comité international des poids et mesures* has noted that isotope delta measurements that cannot presently be made traceable to the International System of Units (SI) should be made traceable to materials recognized as international standards by the IUPAC Commission on Isotopic Abundances and Atomic Weights [7]. Since measurements used to determine the certified carbon isotope delta value in fructose are traceable to internationally recognized standards [3], FRUT-1 serves as a suitable reference material for laboratory quality assurance programs, as outlined in ISO/IEC 17025.

Traceability of carbon isotope delta values in FRUT-1 was established using internationally recognized reference materials. The standard uncertainties for IAEA-CH-6 [8], USGS62 [9], and USGS40 [10] were revised to incorporate the uncertainty associated with the coherence between the reference materials ( $u = 0.029 \text{ ‰}$ ) [6], resulting in the following values:

IAEA-CH-6:  $-10.450(49) \text{ ‰}$     USGS62:  $-14.790(49) \text{ ‰}$     USGS40:  $-26.390(49) \text{ ‰}$

Additionally, IAEA-603, IAEA-610, IAEA-611, and IAEA-612 were used as calibrators. However, the isotope delta values for these materials are certified relative to the VPDB with no reference to LSVEC [11,12]. Consequently, these values were converted to the VPDB scale as defined by IUPAC using both NBS19 and LSVEC [13, 14]. Thus, the following isotope delta values were used for these reference materials:

IAEA-603:  $+2.474(23) \text{ ‰}$                       IAEA-610:  $-9.145(19) \text{ ‰}$   
IAEA-611:  $-30.925(21) \text{ ‰}$                       IAEA-612:  $-36.878(26) \text{ ‰}$

The reassessment of IAEA-603, IAEA-610, IAEA-611, and IAEA-612 did not include the aforementioned additional uncertainty due to lack of coherence [6] because this suite of materials was calibrated independent of the other internationally recognized reference materials.

### **Quality Management System (ISO 17034, ISO/IEC 17025)**

This material was produced in compliance with the NRC Metrology Quality Management System, which conforms to the requirements of ISO 17034 and ISO/IEC 17025. The Metrology Quality Management System supporting NRC Calibration and Measurement Capabilities, as listed in the *Bureau international des poids et mesures* (BIPM) Key Comparison Database ([kcdb.bipm.org/](http://kcdb.bipm.org/)), has been reviewed and approved under the authority of the Inter-American Metrology System (SIM) and found to be in compliance with the expectations of the *Comité international des poids et mesures* (CIPM) Mutual Recognition Arrangement. The SIM approval is available upon request.

### **Updates**

Users should ensure that the certificate they have is current. For updates, please refer to [doi.org/10.4224/crm.2018.frut-1](https://doi.org/10.4224/crm.2018.frut-1)

### **References**

1. JCGM 100:2008. Evaluation of measurement data – Guide to the expression of uncertainty in measurement. [https://www.bipm.org/utis/common/documents/jcgm/JCGM\\_100\\_2008\\_E.pdf](https://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf)

2. Wieser, M.E. (2006) Atomic weights of the elements 2005 (IUPAC Technical Report). *Pure Appl. Chem.* 78(11): 2051-2066. [doi.org/10.1351/pac200678112051](https://doi.org/10.1351/pac200678112051)
3. Brand, W. A., Coplen, T. B., Vogl, J., Rosner, M., and Prohaska, T. (2014). Assessment of international reference materials for isotope-ratio analysis (IUPAC Technical Report). *Pure Appl. Chem.* 86(3): 425-467. [doi.org/10.1515/pac-2013-1023](https://doi.org/10.1515/pac-2013-1023)
4. Chartrand, M.M.G., Meija, J., Hélie, J.-F., Adamowicz-Walczak, A., Middlestead, P., Grinberg, P., Kumkrong, P., Mester, Z. BEET-1: Isotopic Certified Reference Material of beet sugar. Ottawa: National Research Council Canada; 2018. [doi.org/10.4224/crm.2018.beet-1](https://doi.org/10.4224/crm.2018.beet-1)
5. Chartrand, M.M.G., Meija, J., Hélie, J.-F., Adamowicz-Walczak, A., Middlestead, P., Grinberg, P., Kumkrong, P., Mester, Z. GALT-1: Isotopic Certified Reference Material of galactose. Ottawa: National Research Council Canada; 2018. [doi.org/10.4224/crm.2018.galt-1](https://doi.org/10.4224/crm.2018.galt-1)
6. Chartrand, M.M.G., Meija, J., Kumkrong, P. and Mester, Z. (2019) Three sugar certified reference materials for carbon isotope delta ratio measurements. *Rapid Comm. Mass Spectrom.* 33: 272-280. [doi.org/10.1002/rcm.8357](https://doi.org/10.1002/rcm.8357)
7. CIPM. (2015) Decision 104-26 –Traceability exception related to delta value isotope ratio measurements. <https://www.bipm.org/documents/20126/48101963/Traceability-Exception-QM1.pdf/ca63c530-c553-60a1-9111-406c66de4c4d>
8. International Atomic Energy Agency (2007). NBS 22, IAEA-CH-3, IAEA-CH-6, IAEA-CH-7, USGS24 reference sheet. <https://bit.ly/3lfrbXP>
9. US Geological Survey (2019). Report of Stable Isotopic Composition Reference Materials USGS61, USGS62 and USGS63 (Hydrogen, Carbon and Nitrogen Isotopes in Caffeine). <https://www.usgs.gov/media/files/rsil-report-stable-isotopic-composition-reference-materials-usgs61-usgs62-and-usgs63>
10. US Geological Survey (2019). Report of Stable Isotopic Composition Reference Material USGS40 (Carbon and Nitrogen Isotopes in L-glutamic acid). <https://www.usgs.gov/media/files/rsil-report-stable-isotopic-composition-reference-material-usgs40>
11. Assonov, S., Groening, M., Fajgelj, A., Hélie, J.-F., and Hillaire-Marcel, C. (2020). Preparation and characterisation of IAEA-603, a new primary reference material aimed at the VPDB scale realisation for  $\delta^{13}\text{C}$  and  $\delta^{18}\text{O}$  determination. *Rapid Commun. Mass Spectrom.* 34(20): e8867. [doi.org/10.1002/rcm.8867](https://doi.org/10.1002/rcm.8867)
12. Assonov, S., Fajgelj, A., Hélie, J.-F., Allison, C., and Gröning, M. (2021). Characterisation of new reference materials IAEA-610, IAEA-611 and IAEA-612 aimed at the VPDB  $\delta^{13}\text{C}$  scale realisation with small uncertainty. *Rapid Commun. Mass Spectrom.* 35(7): e9014. [doi.org/10.1002/rcm.9014](https://doi.org/10.1002/rcm.9014)
13. Hélie, J.-F., Adamowicz-Walczak, A., Middlestead, P., Chartrand, M. M. G., Mester, Z., and Meija, J. (2021). Discontinuity in the Realization of the Vienna Peedee Belemnite Carbon Isotope Ratio Scale. *Anal. Chem.* 93: 10740-10743. [doi.org/10.1021/acs.analchem.1c02458](https://doi.org/10.1021/acs.analchem.1c02458)
14. Chartrand, M. M. G., Meija, J., Hélie, J.-F., Middlestead, P., Ramalingam, M., Aziz, A. A., and Mester, Z. (2022). Characterization of vanillin carbon isotope delta reference materials. *Anal. Bioanal. Chem.* 414(27): 7877-7883. [doi.org/10.1007/s00216-022-04322-x](https://doi.org/10.1007/s00216-022-04322-x)

### Cited by

A list of scientific publications citing FRUT-1 can be found at [doi.org/10.4224/crm.2018.frut-1](https://doi.org/10.4224/crm.2018.frut-1)

**Authorship**

Michelle M.G. Chartrand<sup>1</sup>, Juris Meija<sup>1</sup>, Jean-Francois Hélie<sup>2</sup>, Agnieszka Adamowicz-Walczak<sup>2</sup>, Paul Middlestead<sup>3</sup>, Patricia Grinberg<sup>1</sup>, Paramee Kumkrong<sup>1</sup>, and Zoltan Mester<sup>1</sup>

<sup>1</sup> National Research Council Canada, 1200 Montreal Rd, Ottawa ON K1A 0R6, Canada

<sup>2</sup> Geotop, Université du Québec à Montréal

<sup>3</sup> Ján Veizer Stable Isotope Laboratory, University of Ottawa

**Acknowledgments**

The contributions of Marie-Pier Thibeault, Kenny Nadeau, Garnet McRae, Jennifer Bates, Enea Pagliano, Lu Yang, and Jeremy Melanson (NRC) are acknowledged.

**Citation**

Chartrand MMG, Meija J, Hélie J-F, Adamowicz-Walczak A, Middlestead P, Grinberg P, Kumkrong P, Mester Z. FRUT-1: Isotopic Certified Reference Material of fructose. Ottawa: National Research Council Canada; 2018.

Available from: [doi.org/10.4224/crm.2018.frut-1](https://doi.org/10.4224/crm.2018.frut-1)

**FRUT-1**

*Date of issue: December 2018*

*Date of expiry: December 2028*

*Revised: January 2020 (editorial updates), November 2023 (expiry date extended, certified carbon isotope delta value revised, editorial updates)*

**Approved by:**

\_\_\_\_\_  
Patricia Grinberg, Ph. D.  
Team Leader, Inorganic Chemical Metrology  
NRC Metrology

**This Certificate is only valid if the corresponding material was obtained directly from the NRC or an Authorized Reseller.**

National Research Council Canada  
Metrology  
1200 Montreal Road  
Building M36, Room 1029  
Ottawa, Ontario K1A 0R6

**Telephone:** 613-993-2359

**Fax:** 613-993-8915

**Email:** [CRM-MRCOttawa@nrc-cnrc.gc.ca](mailto:CRM-MRCOttawa@nrc-cnrc.gc.ca)

